



Product Service

EC-Attestation of Conformity

No. E8 07 02 62688 001

Holder of Certificate: JISICO CO., LTD.

Jisico Building, 301-1, Sungsu-2Ga
133-835 Sungdong-Gu, Seoul
KOREA

Name of Object: Industrial, Scientific and Medical equipment
Separatory Funnel Shaker

Model(s): J-MSFS

Description of Object:

Rated input voltage: 230 Va.c.
Rated frequency: 50-60 Hz
Rated input current: 0.18 A
Rated input power: 60 A

Tested according to:

EN 61326:1997+A1:1998+A2:2001+A3:2003
EN 61000-3-2:2000
EN 61000-3-3:1995+A1:2001

This EC-Attestation of Conformity is issued according to the Directive 89/336/EEC relating to electromagnetic compatibility on a voluntary basis. It confirms that the listed equipment complies with the principal protection requirements of the EMC directive and applies only to the sample and its technical documentation submitted to TÜV SÜD Product Service GmbH for testing and certification. See also notes overleaf.

Test report no.: E072R-022

[Handwritten signatures]



Date, 2007-02-26



After preparation of the necessary technical documentation as well as the conformity declaration the required CE marking can be affixed on the product. Other relevant directives have to be observed.

Page 1 of 1



Product Service

EC-Attestation of Conformity

No. N8 07 04 62688 002

Holder of Certificate: JISICO CO., LTD.

Jisico Building, 301-1, Sungsu-2Ga
133-835 Sungdong-Gu, Seoul
KOREA

Product:

**Electronic measuring equipment
(Separatory Funnel Shaker)**

This EC-Attestation of Conformity is issued on a voluntary basis according to the Low Voltage Directive 2006/95/EC relating to electrical equipment designed for use within certain voltage limits. It confirms that the listed equipment complies with the principal protection requirements of the directive. It refers only to the particular sample submitted for testing and certification. See also notes overleaf.

Test report no.:

CPSA0113467



Date, 2007-04-26



After preparation of the necessary technical documentation as well as the conformity declaration the required CE marking can be affixed on the product. Other relevant directives have to be observed.

Page 1 of 2



Product Service

EC-Attestation of Conformity
No. N8 07 04 62688 002

Model(s): **J-MSFS**

Parameters:

Rated input voltage:	230 V a.c.
Rated frequency:	50/60 Hz
Rated input power:	60 W
Protection class:	I
Installation category:	II
Pollution degree:	2
Degree of protection against ingress of liquids:	IPX0

**Tested
according to:**

EN 61010-2-051:2003
EN 61010-1:2001

ONETECH

Accredited Testing Laboratory under the terms of ISO/IEC 17025

CE CONFORMANCE REPORT

LOW VOLTAGE DIRECTIVE 2006/95/EC

for

Product: Separatory Funnel Shaker

"MODEL NAME(S): J-MSFS"

AUTHORIZED

JISICO CO. LTD.

JISICO B/D, 301-1 2ka, Seongsu-dong,


Seongdong-ku, Seoul, Korea

May 15, 2007

SUMMARY

The equipment complies with the standards;

Safety: EN 61010-1 : 2001 (2nd Edition)



(signature)

G. W. Lee / Managing Director

This is only valid in connection with Test Report: S075R-008

It should not be reproduced except in full, without the written approval of ONETECH Corp.

SAFETY-CE:DoC(Rev.0)

HEAD OFFICE : #505 SK Apt.Factory 223-28 Sangdaewon1-dong, Jungwon-gu, Seongnam-si, Gyeonggi-do, 462-705 Korea
(TEL: 82-31-746-8500 FAX: 82-31-746-8700, URL: <http://www.onetech.co.kr>)