

ONETECH

Accredited Testing Laboratory under the terms of ISO/IEC 17025

CE CONFORMANCE REPORT

EMC DIRECTIVE 89/336/EEC

AMENDED BY DIRECTIVES 92/31/EEC, 93/68/EEC

for

PRODUCT: VORTEX TUBE MIXER

"MODEL NAME: J-MF"

AUTHORIZED

JISICO CO., LTD.


Jisico Building, 301-1, Sungsu-2ga, Sungdong-gu, Seoul, Korea

May 22, 2007

SUMMARY

The equipment complies with the standards:

- EN 61326: 1997 + A1: 1998 + A2: 2001 + A3: 2003 (Class A)
- EN 61000-3-2: 2000
- EN 61000-3-3: 1995 + A1: 2001



(signature)

Y. K. Kwon / Director

This is only valid in connection with Test Report: E075R-043

ONETECH

Accredited Testing Laboratory under the terms of ISO/IEC 17025

CE CONFORMANCE REPORT

LOW VOLTAGE DIRECTIVE 2006/95/EC

for

Product: Vortex Tube Mixer

“MODEL NAME(S): J-MF”

AUTHORIZED

JISICO CO. LTD.

JISICO B/D, 301-1 2ka, Seongsu-dong,

Seongdong-ku, Seoul, Korea


May 31, 2007

SUMMARY

The equipment complies with the standards;

Safety: EN 61010-1 : 2001 (2nd Edition)

EN 61010-2-051 : 2003 (2nd Edition)



(signature)

G. W. Lee / Managing Director

This is only valid in connection with Test Report: S075R-009

It should not be reproduced except in full, without the written approval of ONETECH Corp.

SAFETY-CE:DoC(Rev.0)

HEAD OFFICE : #505 SK Apt.Factory 223-28 Sangdaewon1-dong, Jungwon-gu, Seongnam-si, Gyeonggi-do, 462-705 Korea
(TEL: 82-31-746-8500 FAX: 82-31-746-8700, URL: <http://www.onetech.co.kr>)